

Sakamoto Tube Feeding Trainer II



1. Insertion of catheter
(1) Nasal / oral catheter insertion is possible.



Part Number	Description
IMPM190	Sakamoto Tube Feeding Trainer II
IMPM190-1	Head for IMPM190 Tube Feeding Trainer

Catheters in size range from 5~15Fr can be used.



Mentone Educational Centre
10-12 McWilliam Street Springvale
Victoria, 3171, AUSTRALIA.
Ph: (03) 9547 6638 Fax: (03) 9553 4562
International
Ph: +61 3 9553 3234 Fax: +61 3 9553 4562

Sakamoto Tube Feeding Trainer II



Head moves left and right

(2) It can be set to the appropriate position, such as semi-sitting position.



(3) Using xiphoid process mark as indicator, the length of insertion of the catheter can be measured.



2. Confirmation of reaching of catheter
Reaching of the catheter into the stomach can be confirmed by suctioning gastric juice (water) or auscultating bubble sound after inject air with a syringe.



3. Care of the wearer of the gastric fistula
It is possible to practice caring of the wearer of gastric fistula. Actual nutrient can be injected.



4. Suction
In addition to the suctioning associated to the tube feeding, simple suction function in the intubation tube and tracheal cannula is possible.